

CONVERSION CHART**Leak Rate Units**

	mbar l s ⁻¹	Torr l s ⁻¹	atm cm ³ s ⁻¹	lusec	atm ft ³ min ⁻¹	grammes per year (Freon 12)
1 mbar l s ⁻¹ =	1	0.75	0.987	7.5×10^2	2.097×10^{-3}	1.58×10^5
1 Torr l s ⁻¹ =	1.333	1	1.316×10^3	10^3	2.795×10^{-3}	2.083×10^5
1 atm cm ³ s ⁻¹ =	1.013	0.76	1	7.6×10^2	2.12×10^{-3}	1.582×10^5
1 lusec =	1.333×10^{-3}	0.001	1.32×10^{-3}	1	2.79×10^{-6}	2.083×10^2
1 atm ft ³ min ⁻¹ =	4.78×10^2	3.58×10^2	4.72×10^2	3.58×10^5	1	7.45×10^7
1 gramme per yeay = (Freon 12)	6.4×10^{-6}	4.8×10^{-6}	6.32×10^{-6}	4.8×10^{-3}	1.34×10^{-8}	1

www.avtservices.com.au

1800 55 99 88